

Introduction

Article 48(7) and (9) of Regulation (EU) 2017/746 on in vitro diagnostic medical devices (IVDR) establish the requirement to assess technical documentation of at least one representative device per generic device group (for Class C) and for each category of devices (for Class B) prior to issuing a certificate to a manufacturer.

Section 2.3 and 3.4 of Annex IX define that the quality management system assessment has to be accompanied by assessment of the technical documentation for devices **selected on a representative basis**.

Section 4.5.2(a) of Annex VII requires the Notified Body to draw up and keep up to date, a sampling plan for the assessment of Technical Documentation as referred to in Annexes II and III prior to the audit.

In December 2019 the Medical Device Coordination Group (**MDCG**) published the document "MDCG 2019-13 Guidance on sampling of MDR Class IIa / Class IIb and IVDR Class B / Class C devices for the assessment of the technical documentation" to (amongst others) define the requirements of sampling for Class B and Class C devices under the IVDR for the purpose of assessing the technical documentation.

This document also interprets Section 4.5.2(b) of Annex VII to require the Notified Body to assess the Technical Documentation as preparation for the audit(s). This assessment is expected to be finalised in due time of such audit(s).

This document describes how the above requirements are implemented by **TÜV SÜD Product Service GmbH** and gives additional practical explanations/implications.

Which devices are sampled?

Technical Documentations of non-sterile Class A devices must be kept up-to-date by the legal manufacturer, however, no Notified Body is involved in its assessment. Technical Documentations of Class D, Companion

Diagnostic, Self-Testing and Near-Patient Testing devices are assessed without exception before they can be placed on the market (although not the correct term, this sometimes is called "PMA", pre-market approval).

That leaves Class B and Class C devices, here the assessment of the Technical Documentation of each device is not necessary, but devices are assessed on a representative basis, i.e. "sampled". A "device" should be understood as the device(s) associated with one Basic UDI-DI.

What does sampling mean in this context?

TÜV SÜD needs a complete list of all your devices that are within the scope of the IVDR, including classification and categorisation / grouping.

A Sampling Plan is created including, for each Device Category and Generic Device Group, a schedule defining which devices' Technical Documentations will be assessed when in the course of the certification cycle. The plan is created for 5 years initially and may change whenever there is a change to the device portfolio or due to other input.

Contrary to the common understanding of the term "sampling" and the picture on Page 1 of this document, the process is not completely random, but risk-driven. I.e. TÜV SÜD will create the sampling plan on a risk-based approach, as far as possible.

What is a Device Category and a Generic Device Group?

B

DEVICE CATEGORY

IVR code according to Regulation (EU) 2017/2185



GENERIC DEVICE GROUP

3rd level of European Nomenclature on Medical Devices (EMDN, former Italian CND), consisting of 1 letter and 4 digits plus most appropriate IVP code

Examples for Device Categories (Class B, IVR code)

Excerpt from COMMISSION IMPLEMENTING REGULATION (EU) 2017/2185 of 23 November 2017:

| IVR 0605 | Devices intended to be used for monitoring of levels of medicinal products, substances or biological components | | |
|----------|---|--|--|
| IVR 0606 | Devices intended to be used for non-infectious disease staging | | |
| IVR 0607 | Devices intended to be used for detection of pregnancy or fertility testing | | |
| IVR 0608 | Devices intended to be used for screening, determination or monitoring of physiological markers | | |
| IVR 0609 | Other devices intended to be used to define or monitor physiological status and therapeutic measures | | |
| IVR 0701 | Devices which are controls without a quantitative assigned value | | |
| IVR 0702 | Devices which are controls without a qualitative value | | |

2020 © TÜV SÜD Product Service GmbH | PS-MKT/MHS/19.0/en/SG

Examples for Generic Device Groups (Class C, EDMN + IVP code)

| EDMN CODE | EDMN CODE TITLE | IVP CODE | IVP CODE TOPIC |
|-----------|--|----------|----------------------|
| W0101 | Clinical chemistry | IVP 3002 | Biochemistry |
| W0101 | Clinical chemistry | IVP 3013 | Spectroscopy |
| W0102 | Immunochemsitry (immunology) | IVP 3001 | Agglutination |
| W0102 | Immunochemsitry (immunology) | IVP 3007 | Immunoassays |
| W0103 | Haematology / Haemostasis / Immunoheamotology / Histology / Cytology | IVP 3006 | Flow cytometry |
| W0103 | Haematology / Haemostasis / Immunoheamotology / Histology / Cytology | IVP 3007 | Immunoassays |
| W0104 | Microbiology (culture) | IVP 3014 | Cell function |
| W0105 | Infectious immunology | IVP 3007 | Immunoassays |
| W0105 | Infectious immunology | IVP 3011 | NAT / NGS |
| W0106 | Genetic testing | IVP 3004 | Chromosomal analysis |
| W0106 | Genetic testing | IVP 3011 | NAT / NGS |

How many devices are sampled?



What are the practical implications of sampling, how does the process work at TÜV SÜD Product Service GmbH?

The sampling process works hand in hand with the general certification process, please see also the respective information material.

If you choose to get a conformity assessment according to Annex IX of the IVDR, we will need a list of all your IVD devices, including classification and group /category assignment. Please be aware that the number of groups / categories has a direct impact on the number of devices to be sampled.

TÜV SÜD will evaluate the list and agree or disagree with the classification and grouping/categorisation. In case

there are any discrepancies, your personal Client Handler will contact you and try to resolve them.

We will then create a sampling plan for 5 years and ask you to send the Technical Documentations for the sampled devices. In case there are changes in your product portfolio or design, the plan might be adjusted to reflect these.

There is a separate document that describes the Technical Documentation Submission Requirements, please ask your personal Client Handler.